

**OIML Member State**  
The Netherlands

Number R76/2006-NL1-17.56  
Project number 1901032  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Bizerba SE & Co. KG Wilhelm-Kraut-Str. 65 72336 Balingen Germany
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type: WM-ITC...
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **(III)** or **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
20 November 2017



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-1901032-01 dated 20 November 2017 that includes 58 pages;
- No. NMI-1901032-02 dated 20 November 2017 that includes 47 pages;
- No. NMI-1901032-03 dated 20 November 2017 that includes 33 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	Ⓐ or Ⓑ
Maximum capacity	Max ≤ 60 kg
Verification scale interval	e ≥ 1 g
Weighing range(s)	Single interval Multiple range
Maximum number of scale intervals (one weighing range)	n ≤ 6000 divisions
Maximum number of scale intervals (multiple range)	n ≤ 3200 divisions (per weighing range)
Maximum number of weighing ranges	2
Tare	T ≤ -Max for instruments with one weighing range T ≤ -Max <sub>i</sub> for each range for multiple range instruments
Temperature range	-10 °C / + 40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz
Software identification	Version number: ECn.07.xx (x = 0 ... 99) Software-ID: sid.3b45 (hexadecimal format / upper or lower case letters are possible).