



OIML Certificate of Conformity

OIML Member State
The Netherlands

Number R76/2006-NL1-17.60
Project number 1901707
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Mettler-Toledo (Albstadt) GmbH Unter dem Malesfelsen 34 D-72458 Albstadt Germany
Identification of the certified type	A Terminal / Indicator Type : ICS4...(x), ICS6...
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **II** or **III** or **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
27 November 2017


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R76/2006-NL1-10.25 dated 18 November 2010 that includes 49 pages;
- No. R76/2006-NL1-10.43a dated 23 November 2010 that includes 18 pages;
- No. R76/2006-NL1-10.43b dated 23 November 2010 that includes 28 pages;
- No. NMI-11200439-05 dated 8 March 2012 that includes 19 pages;
- No. NMI-11200439-07 dated 8 March 2012 that includes 25 pages;
- No. NMI-13200233-01 dated 24 October 2013 that includes 19 pages;
- No. NMI-1901294-01 dated 16 October 2017 that includes 12 pages.

Characteristics of the indicator:

Configuration	Digital load cells or weighing module		
Accuracy class OIML R 76	II	III	III
Maximum number of verification scale intervals	100000	10000	1000
Weighing range(s)	Single interval Multi-interval Multiple range		
Power supply voltage	100-230 V AC 50/60 Hz; 9 - 28 V DC through an external AC/DC adapter; 12 V DC through built in battery; For ICS466x and ICS426x 5,3V DC to 12,6V DC in 6 separate excitations.		
Maximum number of load platforms	ICS6x9-1 / ICS4x5-1 / ICS466x: 2 ICS4x9-1, ICS426x: 1 ICS6x5-1: 4		
Temperature range	-10 °C / +40 °C		
Software identification	AA-BB-01.dd.ee-FF-G "AA", "BB", "FF" and "G" can be alphanumeric or numerical characters which describe the configuration like language, application etc., "dd.ee" is the status of the non relevant software part, and "01" is the legally relevant software identification.		