

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-18.01 revision 1  
Project number 2603790  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and  
Manufacturer

Dibal S.A.  
Astintze Kalea, 26 - Pol. Ind. Neinver  
48160 Derio, Vizcaya  
Spain

Identification of the  
certified type

A **Non-automatic weighing instrument**  
Type : KS-JPOS, K-POS

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
19 August 2021

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-18.01 revision 1  
Project number 2603790  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-1901393-01 dated 12 January 2018 that includes 57 pages.
- No. NMI-2603790-01 dated 19 August 2021 that includes 19 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	6 kg ≤ Max ≤ 30 kg
Verification scale interval	e ≥ 2 g
Weighing range(s)	Single interval Multiple range
Maximum number of partial weighing ranges	2
Maximum number of scale intervals (single interval / multiple range)	n ≤ 3000 divisions (per weighing range)
Maximum number of weighing ranges	2
Tare	T ≤ Max - e for single interval instruments T ≤ Max <sub>1</sub> - e <sub>1</sub> for multiple range instruments
Temperature range	5 °C / 40 °C
Power supply voltage	5 V via USB
Application	Intended to be used for direct sales to the public

### Weighing firmware

Weighing CPU	Version number	WLDIBAL, version 2.0.3
	Checksum	110C

### PC software

Software identification:	Framework	Java	.NET (Win XP or equivalent)	.NET (Win 7 or higher)
	Name	JPOSDIBAL	Net XP	Net
	Version number	version 1.x.xx (x = 0... 9)	version: 3.1.x.x (x = 0... 9)	version: 3.0.x.x (x = 0... 9)
	Checksum	8A7DED6411F67452 4CC47FB81	6529ABA3	2ADDCC7C

### Revision History

Revision	Date	Changes
Initial	2018-01-15	-
1	2021-08-19	Adding PC software versions and checksum