

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.01 revision 2
Project number 3067064
Page 1 of 3

Issuing authority NMI Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer Dibal S.A.
Astintze Kalea, 26 - Pol. Ind. Neinver
48160 Derio, Vizcaya
Spain

Identification of the certified type **A Non-automatic weighing instrument**
Type : KS-JPOS, K-POS, KS-POS, KD-POS

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMI Certin B.V., OIML Issuing Authority NL1
4 February 2022

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-1901393-01 dated 12 January 2018 that includes 57 pages.
- No. NMI-2603790-01 dated 19 August 2021 that includes 19 pages.
- No. NMI-3067064-01 dated 4 February 2022 that includes 17 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III		
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$		
Verification scale interval	$e \geq 2 \text{ g}$		
Weighing range(s)	Single interval Multiple range		
Maximum number of partial weighing ranges	2		
Maximum number of scale intervals (single interval / multiple range)	$n \leq 3000$ divisions (per weighing range)		
Maximum number of weighing ranges	2		
Tare	$T \leq \text{Max} - e$ for single interval instruments $T \leq \text{Max}_1 - e_1$ for multiple range instruments		
Temperature range	$5 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$		
Power supply voltage	5 V via USB		
Application	Intended to be used for direct sales to the public		
Firmware identification	Framework	Java	.NET
	Version number	2.0.3	
	Checksum	110C	-

PC software:

Software identification:	Framework	Java	.NET (Win XP or equivalent)	.NET (Win 7 or higher)
	Name	JPOSDIBAL		Net XP
Version number	version 1.x.xx (x = 0... 9)		version: 3.1.x.x (x = 0... 9)	version: 3.0.x.x (x = 0... 9)
Checksum	8A7DED6411F67452 4CC47FB81		6529ABA3	2ADDCC7C



OIML Member State
The Netherlands

OIML Certificate

Number R76/2006-A-NL1-18.01 revision 2
Project number 3067064
Page 3 of 3

Revision History

Revision	Date	Change(s)
Initial	2018-01-15	-
1	2021-08-19	Adding PC software versions and checksum
2	2022-02-04	Adding Type KD-POS