



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.06
Project number 1901130
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	DIGI SINGAPORE PTE LTD 4 Leng Kee Road #06-01 SIS Building Singapore 159088 Singapore
Identification of the certified type	A Non-automatic weighing instrument Type : SM-6000
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
23 January 2018


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.06
Project number 1901130
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report:

No. NMI-1901130-01 dated 22 January 2018 that includes 63 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III	
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$	
Verification scale interval	$e \geq 2g$	
Weighing ranges	Single interval Multi-interval	
Maximum number of scale intervals (one weighing range)	$n \leq 3000$ divisions	
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions	
Maximum number of partial weighing ranges	2	
Tare	$T \leq -\text{Max}_1$ for multi-interval instruments	
Temperature range	$0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$	
Power supply voltage	220 – 240 V AC 50/60 Hz	
Application	Intended to be used for direct sales to the public	
Software identification	Version number	7.1.1.23
	Checksum	89E66100

The above identified Type (represented by the sample(s) identified in the OIML Test Report) have been found to comply with the additional national requirements established by Canada, included in the MAA Declaration of Mutual Confidence:

- R 76 DoMC-01 rev.0, Additional requirements from Canada;
- R 76 DoMC-02 rev.0, Additional requirements from Canada.