



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.14
Project number 1902097
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Shanghai Teraoka Electronic Co. Ltd. Tinglin Industry Developmental Zone Jin Shan District Shanghai 201505 Peoples Republic of China
Identification of the certified type	A Non-automatic weighing instrument Type : RM-80
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
29 March 2018


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.14
Project number 1902097
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R76/2006-NL1-10.39 dated 27 September 2010 that includes 36 pages;
- No. NMI-11200554-01 dated 23 May 2012 that includes 46 pages;
- No. NMI-11200554-02 dated 23 May 2012 that includes 13 pages;
- No. NMI-11200701-01 dated 14 October 2013 that includes 45 pages;
- No. NMI-11200701-02 dated 14 October 2013 that includes 20 pages;
- No. NMI-13200079-01 dated 23 April 2013 that includes 17 pages;
- No. NMI-1902097-01 dated 29 March 2018 that includes 8 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 6000$ divisions
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \leq -50\%$ for instruments with one weighing range $T \leq -\text{Max}_1 - e_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$
Power supply voltage	100 – 240 V AC 50/60 Hz; 12 V DC Battery
Application	Intended to be used for direct sales to the public
Software identification	Version number 2.xx (x = 03... 99)