



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.21
Project number 1901844
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Mettler-Toledo (Albstadt) GmbH Unter dem Malesfelsen 34 D-72458 Albstadt Germany
Identification of the certified type	An indicator / terminal Type : IND890-1 , IND930
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **II** or **III** or **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
28 May 2018


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.21
Project number 1901844
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-1900621-01 dated 24 May 2017 that includes 31 pages;
- No. NMI-1900621-02 dated 24 May 2017 that includes 14 pages;
- No. NMI-1900621-03 dated 6 June 2017 that includes 26 pages.

Characteristics of the indicator / terminal:

Configuration		Digital load cell(s), weighing module or ADPD	
Accuracy class		II or III or IIII	
Weighing range(s)		Single interval Multi-interval Multiple range	
Load cell power supply		5 V DC	
Fraction of the maximum permissible error		0	
Maximum number of load platforms		4	
Temperature range		-10 °C / +40 °C	
Power supply voltage		100 – 240 V AC 50/60 Hz	
Software identification	Software module	Version	Checksum
	Boot service 'Classic' version	V1.1.3	B645
	Boot service 'Pro' version	V2.0.0	EE8D
	ScaleLock	V1.1.xx (xx is a number between 12 and 99)	-
	Scale Module	V1.1.xx (xx is a number between 12 and 99)	
	Scale server 'Classic' version	V1.y.xx (y is a number between 3 and 9) (xx is a number between 0 and 99)	
	Scale server 'Pro' version	V2.y.xx (y is a number between 1 and 9) (xx is a number between 0 and 99)	