

## OIML Certificate

## **OIML Member State**

The Netherlands

Number R76/2006-A-NL1-18.30 Project number 1902544 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and Xiamen Rongta Technology Co., Ltd. Manufacturer 3F/E Building, Gaoqi Industrial Area,

Gaoqi Beisan Road, Dianqian Huli Xiamen City

**Peoples Republic of China** 

Identification of the A

difficulty - Anon-automatic weighing instrument

certified type : AURORA S1

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority NMi Cert

NMi Certin B.V., OIML Issuing Authority NL' 16 November 2018

/ X /

Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org







## **OIML** Certificate

**OIML Member State**The Netherlands

Number R76/2006-A-NL1-18.30 Project number 1902544 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NIM-LSmt2018-0656 dated 25 April 2018 that includes 44 pages;
- No. NMi-1902544-01 dated 16 November 2018 that includes 28 pages.

## **Characteristics of the non-automatic weighing instrument:**

Accuracy class + + + + + + + + + + +	+ + + + + + + + + + + + + + + + + + + +
Maximum capacity	+ + + + 6 kg ≤ Max ≤ 30 kg
Verification scale interval	e ≥ 2 g
Weighing ranges	Single interval or Multi-interval
Maximum number of scale intervals	n ≤ 3000 divisions (for one weighing range for single interval, or per partial weighing range for multi-interval)
Maximum number of partial weighing ranges	+ + + + + + + + 2 + + + + + + + + +
Tare	$T \le -50\%$ for instruments with one weighing range $T \le -Max_1$ for multi-interval instruments
Temperature range	* * * * * * * * * * * * * * * * * * *
Power supply voltage	110 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification A/D board	Version number: V1.xxx (xxx = 118 999) Checksum: FE42432F
Software identification display software	Version number: V1.x.x (x = 1 9) Checksum: 9B69F092

5