



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.31 revision 1
Project number 1901970
Page 1 of 3

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Mettler-Toledo (Changzhou) Measurement Technology Ltd. 111 West Taihu Road, Xinbei District Changzhou, Jiangsu 213125 Peoples Republic of China
Identification of the certified type	An Analog data processing device Type : LE-DigiCell
Characteristics	See next page

This OIML Certificate is issued under scheme A.


This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
10 September 2018


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.31 revision 1
Project number 1901970
Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-11200439-04 dated 8 March 2012 that includes 20 pages;
- No. NMI-12200333-01 dated 12 October 2012 that includes 21 pages;
- No. NMI-15200100-01 dated 3 July 2015 that includes 8 pages;
- No. NMI-15200100-02 dated 3 July 2015 that includes 12;
- No. NMI-1901970-01 dated 5 July 2018 that includes 44 pages;
- No. NMI-1901970-02 dated 5 July 2018 that includes 17 pages.

Characteristics of the indicator:

Accuracy class	III or IIII	
Weighing range(s)	Single interval Multi-interval Multiple range	
Maximum number of scale intervals	$n \leq 3000$ divisions	
Maximum number of weighing ranges	3	
Load cell excitation voltage	2 V DC	
Load cell power supply voltage	-	
Minimum signal input voltage	$U_{\min} = 0$ mV	
Minimum input voltage per verification scale interval	0,35 μ V	
Minimum load cell resistance	87 Ω	
Maximum load cell resistance	1140 Ω	
Fraction of the maximum permissible error	0,5	
Load cell connection	4-wire or 6-wire (remote sensing)	
Maximum value of the cable length per cross wire section between the analog data processing device and the junction box or load cells	No special cable length In case a 4-wire connection is used the load cells are connected directly without junction box	
Maximum number of load platforms	4	
Temperature range	-10 °C / +40 °C	
Power supply voltage	5,4 - 8 V DC battery	
Software identification	Rainbow core	RB or rb xxxxx (Any version)
	Rainbow Weighing Package	WP or wp 2.2.x
	Rainbow Signal processing	SP or sp or DSP or dsp 2.0.x
	Wireless module	BT-FW W1.y.zz



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.31 revision 1
Project number 1901970
Page 3 of 3

Revision History

This revision replaces the previous version(s).

Revision	Date	Change(s)
Initial	6 July 2018	-
1	10 September 2018	Correction of the software identification and power supply voltage.