

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.32
Project number 1902218
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Shanghai Teraoka Electronic Co. Ltd. Tinglin Industry Developmental Zone Jinshan District Shanghai 201505 People's Republic of China
Identification of the certified type	A Non-automatic weighing instrument Type : RM-60; RM-60H
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
19 July 2018

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.32
Project number 1902218
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. R76/2006-NL1-10.39 dated 27 September 2010 that includes 36 pages;
- No. NMI-13200079-01 dated 23 April 2013 that includes 17 pages;
- No. NMI-1902218-01 dated 13 July 2018 that includes 12 pages;
- No. NMI-1902218-02 dated 13 July 2018 that includes 8 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Maximum number of scale intervals	$n \leq 6000$ divisions $n \leq 3000$ divisions for multi-interval instruments (per partial weighing range)
Maximum partial weighing ranges	2
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Tare	$T \leq -50\%$ for single interval instruments $T \leq -\text{Max}_1$ for multi interval instruments
Weighing range(s)	Single interval Multi-interval
Power supply voltage	220 – 240 V AC 50/60 Hz; 12 V DC Battery
Application	Intended to be used for direct sales to the public
Software identification	Version number: 1.xx for the version with STB-2161 main board; 2.xx for the version with STB-2161-1 main board; where xx is a number between 00 and 99, which represents the non-legally relevant software.