

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.40
Project number 1902511
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Manufacturer	Shanghai Teraoka Electronic Co., Ltd. Tinglin Industry Developmental Zone Jin Shan County Shanghai 201505 P.R. of China
Identification of the certified type	A Non-automatic weighing instrument Type : RM-5800III
Characteristics	See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
9 October 2018



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.40
Project number 1902511
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-15200658-01 dated 15 January 2016 that includes 61 pages;
- No. NMI-16200419-01 dated 16 September 2016 that includes 13 pages;
- No. NMI-16200419-02 dated 16 September 2016 that includes 10 pages.
- No. NMI-1902511-01 dated 9 October 2018 that includes 30 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	3 kg ≤ Max ≤ 30 kg
Verification scale interval	e ≥ 1 g
Weighing ranges	Single interval Multi-interval
Maximum number of scale intervals (single interval / multi-interval)	n ≤ 3000 divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	T ≤ -50% for single interval instrument T ≤ -Max ₁ for multi-interval instrument
Temperature range	-10 °C / +40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz 6-12V DC via plug-in power supply
Application	Intended to be used for direct sales to the public (when equipped with a Point Of Sale)
Software identification	Version number: V01.xx (xx = 00 ... 99)