

OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.45
Project number 2232716
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant and Manufacturer	Teraoka Seiko Co. Ltd. 13-12 Kugahara, 5-Chome Ohta-Ku, Tokyo 146-8580 Japan
Identification of the certified type	A Non-automatic weighing instrument Type : DI-771
Characteristics	See next page

This OIML Certificate is issued under scheme A.


This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class  or 

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.
This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
30 October 2018



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-18.45
Project number 2232716
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-13200275-01 dated 11 August 2014 that includes 48 pages;
- No. NMI-16200476-01 dated 30 September 2016 that includes 19 pages.
- No. NMI-10200371-01 dated 25 March 2011 that includes 35 pages;
- No. NMI-10200371-02 dated 25 March 2011 that includes 28 pages.
- No. NMI-10200372-01 dated 25 March 2011 that includes 35 pages;
- No. NMI-10200372-02 dated 25 March 2011 that includes 33 pages.
- No. NMI-10200372-03 dated 25 March 2011 that includes 28 pages.
- No. NMI-10200373-01 dated 2 May 2011 that includes 35 pages;
- No. NMI-10200373-02 dated 2 May 2011 that includes 28 pages.
- No. NMI-10200373-03 dated 2 May 2011 that includes 28 pages.
- No. NMI-10200373-04 dated 2 May 2011 that includes 34 pages.
- No. NMI-15200656-01 dated 10 November 2015 that includes 31 pages.
- No. R60/2000-NL1-10.28A that includes 43 pages;
- No. R60/2000-NL1-10.28B that includes 43 pages;
- No. R60/2000-NL1-10.28C that includes 38 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 1500 \text{ kg}$
Verification scale interval	$e \geq 0,5 \text{ g}$
Actual scale interval	$e = d$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 6000$ divisions
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	3
Tare	$T \leq -100\%$ for instruments with one weighing range $T \leq -\text{Max}1$ for multi-interval instruments
Temperature range	$0 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$
Power supply voltage	100 – 240 V AC 50/60 Hz
Software identification	Version number: 1.05