

OIML Member State The Netherlands

OIML Certificate



Number R76/2006-A-NL1-18.50 revision 1 Project number 2605020 Page 1 of 3

Issuing authority	NMi Certin B.V. Person responsible: M.Ph.D	. Schmidt 🔶
Applicant and Manufacturer	Teraoka Seiko Co., Ltd. 5-13-12, Kugahara, Ohta-ku 146-8580 Tokyo Japan	1,
ldentification of the certified type	A Non-automatic weighi Type	ng instrument : DPS-5600, DPS-5600M
Characteristics	See next page	

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.





NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl NMi Certin B.V., OIML Issuing Authority NL1 13 April 2021

Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.







OIML Certificate

OIML Member State The Netherlands



Number R76/2006-A-NL1-18.50 revision 1 Project number 2605020 Page 2 of 3

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMi-12200108-03 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-04 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-05 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-06 dated 10 October 2012 that includes 18 pages;
- No. NMi-12200108-13 dated 13 May 2013 that includes 15 pages;
- No. NMi-12200108-01 dated 25 June 2014 that includes 49 pages;
- No. R76/1992-NL1-10.11 revision 1 dated 9 January 2015 that includes 26 pages;
- No. NMi-SO15202422-01 dated 19 August 2015 that includes 9 pages;
- No. NMi-15200354-01 dated 31 August 2015 that includes 17 pages;
- No. NMi-16200470-01 dated 22 March 2017 that includes 27 pages;
- No. NMi-16200470-02 dated 22 March 2017 that includes 27 pages;
- No. NMi-16200470-04 dated 22 March 2017 that includes 47 pages;
- No. NMi-16200470-05 dated 22 March 2017 that includes 12 pages;
- No. NMi-16200470-06 dated 22 March 2017 that includes 17 pages;
- No. NMi-16200470-09 dated 20 November 2017 that includes 11 pages;
- No. NMi-2239501-01 dated 11 December 2018 that includes 17 pages;
- No. NMi-2470453-01 dated 4 September 2020 that includes 17 pages;
- No. NMi-2605020-01 dated 13 April 2021 that includes 6 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		
Maximum capacity		3 kg \leq Max \leq 150 kg
Verification scale interval		$e \ge 1 g$
Weighing ranges		Single interval Multi-interval
Maximum number of scale intervals (one weighing range)		$n \le 3000$ divisions
Maximum number of scale intervals (multi-interval)		n ≤ 3000 divisions (per partial weighing range)
Maximum number of partial weighing ranges		2
Tare		$T \leq$ -Max for instruments with one weighing range $T \leq$ -Max_1 for multi-interval instruments
Temperature range		-10 °C / +40 °C
Power supply voltage		100 – 240 V AC 50/60 Hz
Application		Intended to be used for the making-up of prepackages
Software	Measurement software	Version number: 1.xx or 2.xx or 3.xx (xx = 00 99)
identification	A/D board software	Version numbers: 3.xx (For TPB-3356 A/D-board, xx = 22 99), or 1.xx (For TPB-3772 A/D-board, xx = 00 99)



OIML Certificate



Number R76/2006-A-NL1-18.50 revision 1 Project number 2605020 Page 3 of 3

Revision History

This revision replaces the previous version.

)	Revision	Date	Change(s)	
	Initial	11-12-2018	-	
	1	13-04-2021	Added test report NMi-2470453-01 (new main board) Added test report NMi-2605020-01 (remote desktop software)	