

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.02
Project number 1901762
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and
Manufacturer Mettler-Toledo, LLC
1150 Dearborn Drive
Worthington, OH 43085-6712
United States of America

Identification of the
certified type **A Non-automatic weighing instrument**
Type : ACM780

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)** or **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
21 January 2019


C. Oosterman
Head Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.02
Project number 1901762
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

For the indicator:

- No. 511392A dated 20 April 2006 that includes 28 pages;
- No. 511392B dated 20 April 2006 that includes 25 pages;
- No. 511392C dated 20 April 2006 that includes 14 pages;
- No. 811639A dated 1 October 2009 that includes 14 pages;
- No. 811639B dated 1 October 2009 that includes 14 pages;
- No. 9200661A dated 8 February 2010 that includes 19 pages;
- No. 9200661B dated 8 February 2010 that includes 26 pages;
- No. NMI-11200258-01 dated 21 June 2011 that includes 16 pages;

For the complete non-automatic weighing instrument:

- No. NMI-1901762-01 dated 21 January 2019 that includes 31 pages.

Characteristics of the non-automatic weighing instrument:

Maximum number of scale intervals	specified in the certificate(s) of the digital load cell(s) involved			
Weighing range(s)	Single interval Multi-interval Multiple range			
Power supply voltage	100 - 240 V AC 50/60 Hz			
Maximum number of load platforms	4			
Temperature range	-10 °C / +40 °C			
Electromagnetic environment class	E2			
Software identification ACM780	Version	Identification	Remarks	
	1.xx.yy 2.xx.yy 3.xx.yy 4.xx.yy 5.xx.yy	MCN 1.xx	For non-automatic weighing instruments	
	6.xx.yy	6.xx.yy	no MCN present	
	7.xx.yy	7.xx.yy	no MCN present	
	8.xx.yy	8.xx.yy	no MCN present	
	(xx = 00 ... 99) (yy = 00 ... 99)			
	Software identification DataBridge	Program module	Checksum versions 1.9.1 or higher ⁽¹⁾	Remarks
		MTC2.dll	3227273331	-
DataBridgeLRC.dll		2353268258	-	

(1) These version numbers are the software package version numbers, not the version numbers of the legally relevant program modules.