

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-19.13  
Project number 2339311  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: C. Oosterman

Manufacturer and Applicant Marsden Weighing Machine Group Ltd.  
Unit 1, Genesis Business Park  
Sheffield Road, Rotherham S60 1DX  
United Kingdom

Identification of the certified type **A Non-automatic weighing instrument**  
Type : M-999

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
19 February 2019

  
C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-19.13  
Project number 2339311  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report(s):

- No. NMI-1902209-01 rev.1 dated 17 September 2018 that includes 43 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	III
Maximum capacity	Max $\leq$ 250 kg
Verification scale interval	$e \geq 0,5$ kg
Weighing range(s)	Single interval
Maximum number of scale intervals	$n \leq 500$ divisions
Temperature range	+ 5 °C / + 35 °C
Power supply voltage	3,7 V DC by the battery
Application	To be used only for weighing patients
Software identification	Version number P1.xx (xx = 00 ~ 99)