



OIML Certificate

OIML Member State

The Netherlands



Number R76/2006-A-NL1-19.16 Project number 2173264 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and Manufacturer

Quantronix Inc. 380 South 200 West P.O. Box 929

84025 Farmington, Utah **United States of America**

Identification of the certified type

A Non-automatic weighing instrument

Type : Cubiscan 100-LFT

Cubiscan 100-TLFT

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1

15 March 2019

∕osterman Head Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org







NMi Certin B.V. Thiissewea 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl

www.nmi.nl







The Netherlands

OIML Certificate



Number R76/2006-A-NL1-19.16 Project number 2173264 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:



- No. NMi-14200522-01 dated 3 December 2014 that includes 55 pages;
- No. NMi-2173264-01 dated 13 March 2019 that includes 56 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class				III		
Maximum capacity			Max ≤ 50 kg			
Verification scale interval				e ≥ 20 g		
Weighing range			Single interval			
Maximum number of scale intervals		n ≤ 2500				
Temperature range		-10 °C / +40 °C				
Power supply voltage		100 – 240 V AC 50/60 Hz				
Software identification	Version number		4.xxx (x = 0 9) or 5.xxx (x = 0 9)			









