



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.24
Project number 2353595
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer Shanghai Teraoka Electronic Co., Ltd.
Tinglin Industry Developmental Zone
Jin Shan County
Shanghai 201505
P.R. of China

Identification of the certified type **A Weighing module**
Type : AD2000, RM-5800LL B.

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
10 May 2019


C. Oosterman
Head Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.24
Project number 2353595
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-14200679-01 dated 15 February 2015 that includes 43 pages;
- No. NMI-14200679-02 dated 15 February 2015 that includes 28 pages;
- No. NMI-15200365-01 dated 12 June 2015 that includes 15 pages;
- No. NMI-2353595-01 dated 8 May 2019 that includes 13 pages.

Characteristics of the weighing module:

Accuracy class	(III)	
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$	
Verification scale interval	$e \geq 1 \text{ g}$	
Weighing range(s)	Single interval Multi-interval	
Maximum number of scale intervals (one weighing range)	$n \leq 3000$	
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ (per partial weighing range)	
Maximum number of partial weighing ranges	2	
Fraction of the maximum permissible error	1	
Tare	$T \leq -50\%$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments	
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$	
Power supply voltage	For type AD2000: 12 V DC via RS232 5 V DC via USB	For type RM-5800LL B: 100 – 240 V AC 50/60Hz
Software identification	Version number: V1.xx	