



OIML Certificate

OIML Member State

The Netherlands



Number R76/2006-A-NL1-19.26 Project number 2360473 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant Shimadzu Corporation

1, Nishinokyo-Kuwabaracho

Nakagyo-ku Kyoto 604-8511

Japan

Manufacture SHIMADZU PHILIPPINES MANUFACTURING INC.

Phase 3, Lot15, Block 14-15

Cavite EPZ Rosario

Cavite **Philippines**

Identification of the

certified type

A Non-automatic weighing instrument

Type : ATX...R /ATY...R

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (1)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority NMi Certin B.V., OIML Issuing Authority NL1

23 September 2019

∕osterman

Head Certification Board

NMi Certin B.V. Thiissewea 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org











OIML Certificate

OIML Member State The Netherlands



Number R76/2006-A-NL1-19.26 Project number 2360473 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMi-2360473-01 dated 23 September 2019 that includes 20 pages;
- No. NMi-2360473-02 dated 23 September 2019 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	
Maximum capacity	62 g ≤ Max ≤ 320 g, or 600 ct ≤ Max ≤ 1600 ct
Verification scale interval	e ≥ 1 mg, or e ≥ 0,01 ct
Actual scale interval	e = 10 d
Weighing range	Single interval
Maximum number of scale intervals	n ≤ 320000
Temperature range	+10 °C / +30 °C
Power supply voltage	12 V DC, supplied by AC/DC power supply
Software identification	3.00.xx.yy (x, y = 09)







