

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.28
Project number 2371200
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer Shanghai Teraoka Electronics Co., Ltd.
No.6058 of Nan Ting Road
Ting Ling Town, Jin Shan District
Shanghai 201505
China

Identification of the certified type **A Non-automatic weighing instrument**
Type : SM-120

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMI Certin B.V., OIML Issuing Authority NL1
23 May 2019


C. Oosterman
Head Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.28
Project number 2371200
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-14200153-01 dated 16 May 2014 that includes 39 pages;
- No. NMI-14200358-01 dated 30 July 2014 that includes 19 pages;
- No. NMI-15200265-01 dated 8 May 2015 that includes 25 pages;
- No. NMI-1902238-01 dated 20 April 2018 that includes 12 pages;
- No. NMI-2371200-01 dated 23 May 2019 that includes 24 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	(III)
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 300 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing ranges	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 6000$ divisions
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \leq -50\%$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	220 V / 230 V / 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	Version number: VR01.xx (xx = 00... 99)