

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-19.30  
Project number 2212609  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: C. Oosterman

Applicant and Manufacturer Locosc Ningbo Precision Technology Co. Ltd.  
No. 137 Zhenyong Road  
Yongjing, Ningbo, 315021  
China

Identification of the certified type An **Indicator**  
Type : LP7510 Series

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
28 June 2019

  
C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2212609-01 dated 28 June 2019 that includes 50 pages;
- No. NMI-2212609-02 dated 28 June 2019 that includes 14 pages.

### Characteristics of the indicator:

Configuration	Analog load cells
Accuracy class	III or IIII
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals	$n \leq 10000$
Maximum number of partial weighing ranges	2
Load cell excitation voltage	5 V DC
Minimum signal input voltage	$U_{\min} = 0 \text{ mV}$
Minimum input voltage per verification scale interval	1 $\mu\text{V}$
Minimum load cell resistance	87 $\Omega$
Maximum load cell resistance	1218 $\Omega$
Fraction of the maximum permissible error	0,5
Load cell connection	6-wire (remote sensing) or 4-wire
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	3164 m/mm <sup>2</sup> In case a 4-wire connection is used the load cells are connected directly without junction box
Temperature range	-10 °C / +40 °C
Power supply voltage	100 – 240 V AC 50/60 Hz 6 - 9 V DC battery (not suitable for a road vehicle power supply)
Software identification	Version number: PEOxx (for LED display); PCOxx (for LCD display); (xx is a number between 00 and 99)