



## OIML Certificate

## **OIML Member State**

The Netherlands



Number R76/2006-A-NL1-19.31 Project number 2370321 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterman

Applicant and Manufacturer

Mettler-Toledo (Changzhou) Measurement Technology Ltd.

No. 111, West Taihu Road, Changzhou Xinbei District

213125 Jiangsu

Peoples Republic of China

Identification of the

A Non-automatic weighing instrument

certified type

Type bRite

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)



This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.



Issuing Authority

Thiissewea 11 2629 JA Delft

www.nmi.nl

The Netherlands T +31 88 6362332 certin@nmi.nl

NMi Certin B.V., OIML Issuing Authority NL1

16 July 2019



**Oosterman** Head Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org











## **OIML Member State**

The Netherlands



Number R76/2006-A-NL1-19.31 Project number 2370321 Page 2 of 2

**OIML** Certificate

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. R76/2006-NL1-10.31a dated 20 July 2010 that includes 60 pages;
- No. R76/2006-NL1-10.31b dated 20 July 2010 that includes 25 pages;
- No. NMi-14200252-01 dated 4 June 2014 that includes 10 pages.

## **Characteristics of the non-automatic weighing instrument:**

Accuracy class	III
Maximum capacity	3 kg ≤ Max ≤ 30 kg
Verification scale interval	e ≥ 1 g
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	n ≤ 6000 divisions
Maximum number of scale intervals (multi-interval)	n ≤ 6000 divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Temperature range	-10 °C / +40 °C
Power supply voltage	9 – 12 V DC with 100 – 240 V AC/DC adapter; 6 V lead acid battery (option); 7,2 V NiMH battery (option).
Application	Intended to be used for direct sales to the public
Software identification	Version number: 1.xx, or 2.xx  Where xx represents bugfixes and changes in the non-legally relevant part of the software.



