

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.47
Project number 2400355
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer Interpesaje S.A.
Av. Dorado N0. 85D-55 Oficina exterior 35 – Centro
Empresarial Dorado Plaza
Bogota
Colombia

Identification of the certified type **A Non-automatic weighing instrument**
Type : Truck Scale Series FP-UM,
Series SD-UM and Series FP-CP

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
22 August 2019


C. Oosterman
Head Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.47
Project number 2400355
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

For the load cell:

- R60/2000-NL1-04.16 dated 27 September 2004 that includes 53 pages (load cells);
- NMI-1901068-01 dated 22 August 2017 that includes 67 pages (load cells).

For the indicator:

- DANAK-1919104, dated 28 April 2018, that includes 69 pages (indicator).

For each model, the compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in OIML R76 edition 2006 Annex F.5.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	$25 \text{ t} \leq \text{Max} \leq 120 \text{ t}$
Verification scale interval	$e \geq 10 \text{ kg}$
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	$n \leq 4000$ divisions
Tare	$T \leq -\text{Max}$
Temperature range	$-10 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$
Power supply voltage	90 – 240 V AC 50/60 Hz 10-30 V DC
Software identification	Version number 01.xx (x = 0... 99)