

OIML Member State
The Netherlands

Number R76/2006-A-NL1-19.49
Project number 2305042
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: C. Oosterman

Applicant and Manufacturer Changzhou Sunmi Scale Technology Co.,Ltd.
Xinzha Technology Industrial Park
Buidling 19&21, 213000
Changzhou, Jiangsu
China

Identification of the certified type **A Non-automatic weighing instrument**
Type : S2

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III** or **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
27 September 2019


C. Oosterman
Head Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report:

- No. NMI-2305042-01 dated 27 September 2019 that includes 63 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals	$n \leq 3000$
Maximum number of weighing ranges	2
Tare	$T \leq -50\% + e$ for instruments with one weighing range $T \leq -\text{Max}_1 + e_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	100 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	10xxx (x = 0... 9)
Version number	