



# OIML Certificate

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-19.53  
Project number 2399933  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: C. Oosterman

Applicant and Manufacturer TANITA Corporation  
14-2, 1-Chome, Maeno-Cho, Itabashi-Ku  
Tokyo 174-8630  
Japan

Identification of the certified type A **Non-automatic weighing instrument**  
Type : PW-650A, PW-650MA

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
18 October 2019

  
C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. R76/1992-NL1-09.27A dated 1 October 2009 that includes 34 pages;
- No. R76/1992-NL1-09.27B dated 1 October 2009 that includes 24 pages;
- No. R76/1992-NL1-10.35 dated 26 August 2010 that includes 35 pages;
- No. NMI-2399933-01 dated 18 October 2019 that includes 20 pages;
- No. NMI-2399933-02 dated 18 October 2019 that includes 14 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	III		
Maximum capacity	Max ≤ 200 kg		
Verification scale interval	e ≥ 100 g		
Weighing range(s)	Single interval		
Maximum number of scale intervals	n ≤ 2000		
Tare	T ≤ -50%		
Temperature range	+5 °C / +35 °C		
Power supply voltage	230 V AC 50/60 Hz to 9 V DC adapter Or 6 x 1,5 V battery		
Application	Intended to be used for weighing patients		
Software identification	Display software	PV650	9351
	AD converter software	SC240	9201