

**OIML Member State**  
The Netherlands

Issuing authority NMI Certin B.V.  
Person responsible: F. van Booma-de Smit

Applicant and Manufacturer Mettler Toledo GmbH  
Im Langacher 44  
8606 Greifensee  
Switzerland

Identification of the certified type **A Digital data processing device**  
Type : Powercell option board

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMI Certin B.V., OIML Issuing Authority NL1**  
16 December 2019

Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2373298-01 dated 16 December 2019 that includes 17 pages;
- No. NMI-2373298-02 dated 16 December 2019 that includes 15 pages.

**Characteristics of the Digital data processing device:**

Configuration	Digital load cells or weighing module
Accuracy class	III
Weighing range	Single interval Multi interval Multi range
Load cell power supply voltage	12 or 24 V DC
Power supply of DDPD	12 V DC
Maximum number of scale intervals	10000
Fraction of the maximum permissible error	0
Temperature range	-10 °C / +40 °C
Firmware identification	V1.xx.yyyy (x = 0...9; y = 0...9)

The identification number will be displayed on a terminal or indicator presenting the primary indication after pressing the key sequence:

- Selecting "System" on the main page;
- Selecting "Terminal", then "Metrology".