

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-20.03  
Project number 2441538  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: M. Boudewijns

Applicant and Manufacturer Qingdao Wintec System Co., Ltd.  
Wintec Park, Xinye Road  
High-Tech Zone, Qingdao, 266114  
China

Identification of the certified type **A Non-automatic weighing instrument**  
Type : ACS-Sxxx series

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **(III)** or **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
17 February 2020

Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-20.03  
Project number 2441538  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

For the Analog data processing device:

- Qingdao Wintec System Co., Ltd., type WT-C023:
  - No. NMI-2181352-01 dated 22 July 2019 that includes 37 pages;
  - No. NMI-2181352-02 dated 22 July 2019 that includes 33 pages.

For the Terminal:

- Qingdao Wintec System Co., Ltd., type ACS-Sxxx:
  - No. NMI-2181352-01 dated 22 July 2019 that includes 37 pages.

For the load cells:

- Zhonghang Electronic Measuring Instrument Co., Ltd. (ZEMIC), type L6D-xx-xxx-xxx-xx Series:
  - No. NMI-10200947-07 dated 24 December 2010 that includes 59 pages;
  - No. NMI-10200947-08 dated 24 December 2010 that includes 59 pages.

For each model, the compatibility of load cells and analog data processing device is established by the manufacturer by means of the compatibility of modules form, contained in OIML R76 edition 2006 Annex F.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III or IIII
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals	$n \leq$ number of scale intervals in the certificates involved
Application	Intended to be used for direct sales to the public
Software identification	Version number: 2.0.0.1 Checksum: 1DAF85