

OIML Member State The Netherlands

OIML Certificate



Number R76/2006-A-NL1-20.13 Project number 2399622 Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: M. Boud	dewijr	าร		
Applicant and Manufacturer	Mettler-Toledo (Changzhou No.111 West TaiHu Road Jiangsu, 213125 Peoples Republic of China	ı) Mea	asuren	nent Tech	nology
ldentification of the certified type	A Non-automatic weighi i Type	ng ins	strum :	ent BPA121	

Characteristics

This OIML Certificate is issued under scheme A.


See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)


This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.





NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl



Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.







-

OIML Member State The Netherlands

OIML Certificate



Number R76/2006-A-NL1-20.13 Project number 2399622 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMi-2399622-01 dated 11 March 2020 that includes 45 pages;
- No. NMi-2399622-02 dated 11 March 2020 that includes 28 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class		
Maximum capacity	3 kg ≤ Max ≤ 30 kg	
Verification scale interval	e ≥ 1 g	
Weighing range(s)	Single interval	
Maximum number of scale intervals (single interval)	n ≤ 3000	
Temperature range	-10 °C / +40 °C	
Power supply voltage	9 V DC through plug-in power supply 6 V DC with internal battery	
Software identification	2.yy.xxxx (y, x = 0 9)	

The software identification is displayed at start-up.

The non-automatic weighing instrument has embedded software.