

OIML Member State
The Netherlands

Issuing authority NMI Certin B.V.
Person responsible: M. Boudewijns

Applicant and Manufacturer Fuzhou Lucky Star Co., Ltd.
No.5, Shi Zi Ting Road, Cangshan Park
Pushang Industrial District, Fuzhou, Fujian
China

Identification of the certified type An **Indicator** Type

: XK-T1011PR
XK-T1011PG
XK-T1021SR
XK-T1021SG
XK-T1021FR
XK-T1021FG
XK-T1031SR
XK-T1031SG
XK-T1051PR
XK-T1051PG

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class $\textcircled{\text{III}}$ or $\textcircled{\text{III}}$

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
25 May 2020

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-12200515-01 dated 12 March 2014 that includes 48 pages;
- No. NMI-12200515-02 dated 12 March 2014 that includes 12 pages.

Characteristics of the indicator:

Configuration	Analog load cells
Accuracy class	III or IIII
Weighing range	Single interval
Maximum number of scale intervals	$n \leq 3000$
Load cell excitation voltage	5 V DC
Minimum signal input voltage	$U_{\min} = 0 \text{ mV}$
Minimum input voltage per verification scale interval	1 μV
Minimum load cell resistance	87 Ω
Maximum load cell resistance	1220 Ω
Fraction of the maximum permissible error	0,5
Load cell connection	4-wire
Maximum value of the cable length per cross wire section between the indicator and the junction box or load cells	The load cells are connected directly without junction box
Temperature range	-10 °C / +40 °C
Power supply voltage	7,5 V DC supplied by an AC/DC adapter 6 V DC supplied by a battery
Software identification	Version number: VEr21.3