

OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.26
Project number 2450950
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: M. Boudewijns

Applicant and Manufacturer Shanghai Teraoka Electronic Co., Ltd
Tinglin Industry Developmental Zone
Jinshan District,
Shanghai 201505
P.R. of China

Identification of the certified type **A Non-automatic weighing instrument**
Type : DS-867N

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMI Certin B.V., OIML Issuing Authority NL1
15 May 2020

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-15200658-01 dated 15 January 2016 that includes 61 pages;
- No. NMI-16200419-01 dated 20 September 2016 that includes 13 pages;
- No. NMI-2450950-01 dated 14 May 2020 that includes 14 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ divisions (per partial weighing range)
Maximum number of partial weighing ranges	2
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	9 - 12 V DC supplied by an AC/DC adapter
Application	Intended to be used for direct sales to the public
Software identification	Version number: V01.xx (xx = 00...99)

The software identification is displayed at start-up.

When an EPOS is present and it is able to show the primary indications, the non-automatic weighing instrument display may be absent.