



OIML Member State

The Netherlands

Applicant and

Manufacturer



Number R76/2006-A-NL1-20.39 revision 1 Project number 2582084 Page 1 of 4

OIML Certificate

Issuing authority NMi Certin B.V.

Person responsible: M.Ph.D. Schmidt

Shanghai Teraoka Electronic Co., Ltd.

Tinglin Industry Developmental Zone

Jin Shan County Shanghai 201505 P.R. of China

Identification of the

certified type

A Non-automatic weighing instrument

RM-5800 series; Type

SM-120 (LL) BS PLUS

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class



This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.



Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1 11 May 2021



Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.







NMi Certin B.V. Thiisseweg 11 2629 JA Delft The Netherlands T+31 88 6362332 certin@nmi.nl www.nmi.nl







OIML Member State

The Netherlands



Number R76/2006-A-NL1-20.39 revision 1 Project number 2582084 Page 2 of 4

OIML Certificate

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. R76/2006-NL1-10.06 dated 22 March 2010 that includes 59 pages;
- No. NMi-12200263-01 dated 28 September 2012 that includes 23 pages;
- No. NMi-12200684-01 dated 6 December 2012 that includes 11 pages;
- No. NMi-13200468-01 dated 21 October 2013 that includes 21 pages;
- No. NMi-14200357-01 dated 10 October 2014 that includes 37 pages;
- No. NMi-14200490-01 dated 10 October 2014 that includes 17 pages;
- No. NMi-15200739-01 dated 24 February 2016 that includes 35 pages;
- No. NMi-1900734-01 dated 20 April 2017 that includes 42 pages;
- No. NMi-1901725-01 dated 24 January 2018 that includes 15 pages;
- No. NMi-1902212-01 dated 25 May 2018 that includes 28 pages;
- No. NMi-1902212-02 dated 25 May 2018 that includes 25 pages;
- No. NMi-2168906-01 dated 12 June 2019 that includes 47 pages;
- No. NMi-2326018-01 dated 14 June 2019 that includes 13 pages;
- No. NMi-2406858-01 dated 31 October 2019 that includes 12 pages;
- No. NMi-2480079-01 dated 10 July 2020 that includes 16 pages;
- No. NMi-2487322-01 dated 6 August 2020 that includes 58 pages;
- No. NMi-2497211-01 dated 14 September 2020 that includes 21 pages;
- No. NMi-2520358-01 dated 29 October 2020 that includes 15 pages;
 No. NMi-2582084-01 dated 11 May 2021 that includes 16 pages.















OIML Certificate



Number R76/2006-A-NL1-20.39 revision 1 Project number 2582084 Page 3 of 4

Characteristics of the non-automatic weighing instrument:

Load cell us	sed			M-type				
Accuracy class		K-type M-type						
Maximum capacity		3 kg ≤ Max ≤ 30 kg)	30 kg ≤ Max ≤ 600 kg			
Verification scale interval		e ≥ 0,5 g		e ≥ 10 g				
Weighing ranges			Single in	nterval or Multi-interval				
Maximum number of scale intervals		$n \le 6000$ divisions (single interval, with A/D board STB-2236), or $n \le 3000$ divisions (per partial weighing range)						
Maximum partial weighing ranges		2						
Tare		$T \le -50\%$ for instruments with one weighing range $T \le -Max_1$ for multi-interval instruments						
Temperature range		-10 °C / +40 °C						
Power supply voltage		100 – 240 V AC 50/60 Hz						
Application		Integrated POS software Intended to be used for direct sales to the public						
Software identification	on	Filenames	Vers	ion	Checksum		Remarks	
Scale software		libwm.jar	1.1.0.release 1.1.0.release 1 (Maxlabel	e_build:51	00085704 00077148 14015C4F		-	
		libscale.jar libprice.jar	1.1.0.release 1.1.0.release	_	00064855			
POS software	Windows	libscale.dll libprice.dll	1 1	(+)	004591	69 or 4104	17A60	one checksum calculated over two files
	Linux	libscale.so libprice.so	1		007479	97 or 1693	30BA	
		Rm5800drv.so	1.0b	145		20200D92I 2A4D099F		-











OIML Member State The Netherlands



Number R76/2006-A-NL1-20.39 revision 1 Project number 2582084 Page 4 of 4

Certificate history: This revision replaces the previous version.



Revision	Date	Description of the modification
Initial	10 July 2020	-
1	11 May 2021	Adding test report for new mainboard GF40









