

OIML Member State
The Netherlands

Issuing authority NMI Certin B.V.
Person responsible: M. Boudewijns

Applicant and Manufacturer Kern & Sohn GmbH
Ziegelei 1
D-72336 Balingen-Frommern
Germany

Identification of the certified type **A Non-automatic weighing instrument**
Type : TMBA...-A
TMBA...-B

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMI Certin B.V., OIML Issuing Authority NL1
27 August 2020

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report:

- No. NMI-2408004-01 dated 27 August 2020 that includes 53 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$15 \text{ kg} \leq \text{Max} \leq 25 \text{ kg}$
Verification scale interval	$e \geq 5 \text{ g}$
Weighing range	Single interval
Maximum number of scale intervals	$n \leq 3000$
Tare	$T \leq -\text{Max} + e$
Temperature range	+10 °C / +40 °C
Power supply voltage	6,5 – 13 V DC by AC/DC adapter 6 x 1,5 V DC by AA battery
Application	Intended to be used for weighing patients
Software identification	Legally relevant version number A.1.0.10
	Non-legally relevant version number B.x.x.xx (xx = 00 .. 99)

The legally relevant software identification is displayed after pressing the key sequence:

- Press any button at start-up during the display check.