

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-20.47  
Project number 2496179  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: M. Boudewijns

Applicant and Manufacturer Xiamen Pinnacle Electrical Co. Ltd.  
4F GuangXia Building  
Torch High-Tech Zone Xiamen  
CN 361006 FUJIAN  
P.R. of China

Identification of the certified type **A Non-automatic weighing instrument**  
Type : LS2Z, LS5, CS5, LS4 series

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
1 September 2020

Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-15200660-01 dated 14 January 2016 that includes 33 pages;
- No. NMI-15200660-02 dated 14 January 2016 that includes 47 pages.
- No. NMI-2496179-01 dated 1 September 2020 that includes 14 pages.

**Characteristics of the non-automatic weighing instrument:**

Accuracy class	III	
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 75 \text{ kg}$	
Verification scale interval	$e \geq 1 \text{ g}$	
Weighing range(s)	Single interval Multi-interval	
Maximum number of scale intervals	$n \leq 3000$	
Maximum number of partial weighing ranges	2	
Tare	$T \leq -50\%$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments	
Temperature range	$0 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$	
Power supply voltage	100 – 240 V AC 50/60 Hz through AC/DC adapter	
Application	Intended to be used for direct sales to the public Intended to be used for the making-up of prepackages	
Software identification	Version number	V6.058
		V7.360