

OIML Member State The Netherlands

Issuing authority

OIML Certificate



Number R76/2006-A-NL1-20.50 Project number 2532570 Page 1 of 2

| | Person responsible: M. Boudewijns | |
|--------------------------------------|---|-------------------------------|
| Manufacturer | Dibal S.A. Astintze 26 48160 Derio Spain | |
| Identification of the certified type | A Non-automatic weigh Type | ing instrument : CS Series |
| Characteristics | See next page | |

NMi Certin B.V.


This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.





NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl NMi Certin B.V., OIML Issuing Authority NL1 30 December 2020

Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.







OIML Member State The Netherlands

OIML Certificate



Number R76/2006-A-NL1-20.50 Project number 2532570 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMi-1902587-01 dated 14 December 2018 that includes 53 pages;
- No. NMi-1902587-02 dated 14 December 2018 that includes 12 pages;
- No. NMi-1902587-03 dated 14 December 2018 that includes 12 pages;
- No. NMi-2431310-01 dated 18 March 2020 that includes 12 pages;
- No. NMi-2532570-01 dated 30 December 2020 that includes 13 pages.

Characteristics of the non-automatic weighing instrument:

| Accuracy class | | | |
|---|-----------------------|--|--|
| Maximum capacity | | 6 kg ≤ Max ≤ 100 kg | |
| Verification scale interval | | e ≥ 2 g | |
| Weighing ranges | | Single interval Multi-interval Multiple range | |
| Maximum number of scale intervals (one weighing range) | | n ≤ 6000 | |
| Maximum number of scale intervals (multi-interval) | | $n \le 3000$ (per partial weighing range) | |
| Maximum number of the partial weighing ranges | | 2 | |
| Maximum number of scale intervals (multiple range) | | $n \le 3000$ (per weighing range) | |
| Maximum number of weighing ranges | | 2 | |
| Tare | | Tare \leq -Max + e for single interval Tare \leq -Max ₁ + e ₁ for multi-interval and multiple range instruments. | |
| Temperature range | | 0 °C / 40 °C | |
| Power supply voltage | | 110 – 230 V AC 50/60 Hz | |
| Application | | Intended to be used for direct sales to the public | |
| Software identification | Weighing CPU Firmware | Version: 2.0.3 | |
| | PC Software checksums | PHP: F244B8F0 | |
| | | JPOS: ADD09326 | |
| | | Net: 2ADDCC7C | |