

OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.62
Project number 2532564
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: M. Boudewijns

Applicant and Manufacturer Dibal S.A.
Astintze Kalea 26
48160 Derio
Spain

Identification of the certified type **A Non-automatic weighing instrument**
Type : KS-400

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
4 February 2021

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



OIML Member State
The Netherlands

Number R76/2006-A-NL1-20.62
Project number 2532564
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

- No. NMI-16200563-01 dated 14 March 2017 that includes 60 pages;
- No. NMI-16200563-02 dated 14 March 2017 that includes 32 pages.
- No. NMI-2532564-01 dated 4 February 2021 that includes 20 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 2 \text{ g}$
Weighing ranges	Single interval Multiple range
Maximum number of partial weighing ranges	2
Maximum number of scale intervals (single interval / multiple range)	$n \leq 3000$ divisions (per weighing range)
Maximum number of weighing ranges	2
Tare	$T \leq \text{Max} - e$ for single interval instruments $T \leq \text{Max}_1 - e_1$ for multiple range instruments
Temperature range	+ 5 °C / 40 °C
Power supply voltage	5-12 V DC via plug in power supply 5 V via USB 100 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public
Software identification	Legally relevant software version number: V-1.01