



OIML Certificate

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.10
Project number 2446349
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: M. Boudewijns

Applicant and Manufacturer InBody Co., Ltd.
15, Heugam-gil, Ippang-myeon, Seobuk-gu
Cheonan-si
Chungcheongnam-do 31025
South Korea

Identification of the certified type **A Non-automatic weighing instrument**
Type : InBody970

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMI Certin B.V., OIML Issuing Authority NL1
26 February 2021

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report:

- No. NMI-2446349-01 dated 26 February 2021 that includes 54 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	300 kg
Verification scale interval	e = 100 g
Weighing range	Single interval
Maximum number of scale intervals	n ≤ 3000
Temperature range	+10 °C / +40 °C
Power supply voltage	12 V DC from 100-240 V AC 50/60 Hz plugin power supply
Application	Intended to be used for weighing patients
Software identification	IB970-1.x (x = 0... 9 and represents the non-legally relevant software)