

OIML Member State
The Netherlands

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

**Applicant and
Manufacturer**

Bizerba SE & Co. KG
Wilhelm-Kraut-Str. 65
72336 Balingen
Germany

Identification of the
certified type

A **Non-automatic weighing instrument**
Type : Q1 series

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
15 April 2021

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMI-2477454-01 dated 15 April 2021 that includes 51 pages;
- No. NMI-2477454-02 dated 15 April 2021 that includes 18 pages;
- No. NMI-2529842-01 dated 15 April 2021 that includes 17 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max ≤ 30 kg
Verification scale interval	e ≥ 1 g
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals	n ≤ 6000
Maximum number of scale intervals (multi-interval)	n ≤ 3000 (per partial weighing range)
Maximum number of partial weighing ranges	2
Temperature range	-10 °C / +40 °C
Power supply voltage	120 – 240 V AC 50/60 Hz
Application	Intended to be used for direct sales to the public

Terminal software identification:

	Software version number (legally relevant) ¹⁾	Software ID (Windows)	Software ID (Linux)
Scale OEM Module	013	7640	0369
	015	8282	1171

A/D board software identification:

	Version number (legally relevant) ¹⁾	Software ID
ADW	010	6153
	011	5232

¹⁾ The identification of the non-legally relevant software version can optionally be added to the identification of the legally relevant software version, separated by "::" (example: "010::x.y" or "015::x.y").