

OIML Member State
The Netherlands

Number R76/2006-A-NL1-21.25
Project number 2551350
Page 1 of 2

Issuing authority NMI Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer Yamato Scale Co., Ltd.
5-22 Saenba-cho
Akashi
Japan

Identification of the certified type A **Non-automatic weighing instrument**
Type : DP-6701 series

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
7 June 2021

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report:

- No. NMI-2551350-01 dated 7 June 2021 that includes 58 pages.

Characteristics of the non-automatic weighing instrument:

Accuracy class	(III)	
Maximum capacity	30 kg	$32 \text{ kg} \leq \text{Max} \leq 150 \text{ kg}$
Verification scale interval	$e \geq 0,01 \text{ kg}$	
Weighing range	Single interval	
Maximum number of scale intervals (one weighing range)	$n \leq 3000$	$n \leq 3200$
Tare	$T \leq -\text{Max}$	
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$	
Power supply voltage	3,7 V – 6 V DC by non-rechargeable battery, or 100 – 240 V AC/DC to 5,9 V DC by plug-in adapter	
Software identification	Version number	V1.xx. (x = 0... 99)

The software identification is displayed at start-up.