

OIML Member State
The Netherlands

Issuing authority

NMi Certin B.V.
Person responsible: M.Ph.D. Schmidt

Applicant and
Manufacturer

GLOBAL TRADE INDUSTRY
7, Rue Berlioze - Belvedere
20300 Casablanca
Morroco

Identification of the
certified type

An **Indicator**
Type

: DIS1000

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 2006 for accuracy class **(III)**, **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1
14 June 2021

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMI-11200809-08 dated 7 December 2012 that includes 48 pages.

Characteristics of the indicator:

Accuracy class	III	III
Maximum number of verification scale intervals	$n \leq 3000$	$n \leq 1000$
Load cell excitation voltage	5 V DC	
Minimum input voltage per verification scale interval	1,5 μ V	
Minimum load cell resistance	43,8 Ω	
Maximum load cell resistance	1050 Ω	
Temperature range	-10 $^{\circ}$ C / +40 $^{\circ}$ C	
Fraction of the maximum permissible error	0,5	
Load cell connection	6-wire (remote sensing)	
Maximum value of the cable length per cross wire section (6-wire system)	No special cable length has to be provided for the connection between the indicator and the junction box or load cells.	
Weighing range(s)	Single interval	
Power supply voltage	110 – 240 V AC 50/60 Hz	
Software identification	Version number: v0.1	