



**OIML Member State**  
The Netherlands

# OIML Certificate

Number R76/2006-A-NL1-21.37  
Project number 2658948  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer Teraoka Seiko Co., Ltd.  
5-13-12, Kugahara, Ohta-ku,  
146-8580 Tokyo  
Japan

Identification of the certified type A **Non-automatic weighing instrument**  
Type : RGW-560S

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMI Certin B.V., OIML Issuing Authority NL1**  
30 September 2021

Certification Board

NMI Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-21.37  
Project number 2658948  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-12200108-01 dated 25 June 2014 that includes 49 pages;
- No. NMI-15200354-01 dated 31 August 2015 that includes 17 pages;
- No. NMI-16200470-01 dated 22 March 2017 that includes 27 pages;
- No. NMI-16200470-02 dated 22 March 2017 that includes 27 pages;
- No. NMI-16200470-04 dated 22 March 2017 that includes 47 pages;
- No. NMI-16200470-06 dated 22 March 2017 that includes 17 pages;
- No. NMI-16200470-10 dated 20 October 2017 that includes 15 pages;
- No. NMI-2239501-01 dated 20 November 2018 that includes 17 pages;
- No. NMI-2470453-01 dated 4 September 2020 that includes 15 pages;
- No. NMI-2605020-01 dated 14 April 2020 that includes 6 pages.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval Multi-interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$
Maximum number of scale intervals (multi-interval)	$n \leq 3000$ (per partial weighing range)
Maximum number of partial weighing ranges	2
Tare	$T \leq -\text{Max}$ for instruments with one weighing range $T \leq -\text{Max}_1$ for multi-interval instruments
Temperature range	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Power supply voltage	220 – 240 V AC 50/60 Hz
Application	Intended to be used for the making-up of prepackages

The software identification can be recalled from the operation mode:

- Go to the main menu;
- Press the arrow button (in the bottom right corner of the screen);
- Choose setup;
- Choose software version;
- Scroll to the right module name.

### Program identification:

Program module	Version	Remarks
measurement software	1.xx *)	Not suitable for direct sales to the public
	2.xx *)	
	3.xx *)	
	4.xx *)	
		Suitable for direct sales to the public

\*) "xx" is a number between 00 and 99 that represents minor versions that contain bug fixes and non-legally relevant changes