

**OIML Member State**  
The Netherlands

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

Applicant

Mettler-Toledo GmbH  
Im Langacher 44  
CH-8606 Greifensee  
Switzerland

Manufacturer

Mettler-Toledo (Changzhou) Measurement Technology Co., Ltd.  
No:111 West Taihu Road  
Changzhou City  
Jiangsu Province  
P.R. of China

Identification of the certified type

**A Non-automatic weighing instrument**  
Type : BBA23x...

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
28 December 2021

Certification Board

NMi Certin B.V.  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-21.42  
Project number 2631963  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Reports:

Indicator type IND231, IND236, IND236-TLF:

- No. NMI-13200101-01 dated 22 April 2013 that includes 49 pages;
- No. NMI-13200101-02 dated 22 April 2013 that includes 13 pages;
- No. NMI-14200515-01 dated 28 November 2014 that includes 12 pages;

Analog load cell type 0785:

- No. R60/2000-NL-01.06A dated 15 February 2001, that includes 40 pages;
- No. R60/2000-NL-01.06B dated 15 February 2001, that includes 37 pages;
- No. R60/2000-NL1-04.15A dated 30 September 2004, that includes 40 pages;
- No. R60/2000-NL1-04.15B dated 30 September 2004, that includes 37 pages;

Analog load cell type 0795:

- No. R60/2000-NL-01.03 dated 19 January 2001, that includes 40 pages;
- No. NMI-2555505-01 dated 19 March 2021 that includes 16 pages;

Analog load cell type 0805:

- No. R60/2000-NL-02.17A dated 17 June 2002, that includes 43 pages;
- No. R60/2000-NL-02.17B dated 6 June 2002, that includes 37 pages;
- No. R60/2000-NL1-04.19A dated 21 December 2004, that includes 43 pages;
- No. R60/2000-NL1-04.19A revision 1 dated 13 February 2012 that includes 31 pages;
- No. R60/2000-NL1-04.19B dated 21 December 2004, that includes 37 pages;

Analog load cell type SLP84x:

- No. NMI-2495659-04 dated 26 April 2021 that includes 23 pages;
- No. NMI-2495659-05 dated 26 April 2021 that includes 45 pages;
- No. NMI-2495659-06 dated 26 April 2021 that includes 32 pages.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4, at the time of putting into use.

### Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	$3 \text{ kg} \leq \text{Max} \leq 600 \text{ kg}$
Verification scale interval	$e \geq 1 \text{ g}$
Weighing range(s)	Single interval
Maximum number of scale intervals (one weighing range)	$n \leq 3000$
Temperature range	$-10 \text{ }^\circ\text{C} / 40 \text{ }^\circ\text{C}$
Power supply voltage	100-240 V AC 50/60 Hz, or 9 V DC by battery
Software identification	Version number: 1.xx.yyyy (x,y=0...9)

Software:

- The identification number will be displayed at start-up.