



OIML Certificate

OIML Member State The Netherlands



Number R76/2006-A-NL1-22.09 Project number 2733270 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: M.Ph.D. Schmidt

Teraoka Seiko Co.,Ltd. 5-13-12 Kugahara, Ohta-Ku

Tokyo 146-8580

Japan

Identification of the

certified type

Applicant and

Manufacturer

A Non-automatic weighing instrument

SBS-1000

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.



Issuing Authority

NMi Certin B.V., OIML Issuing Authority NL1 28 February 2022



NMi Certin B.V. Thiissewea 11 2629 JA Delft The Netherlands T +31 88 6362332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.











OIML Certificate

Number R76/2006-A-NL1-22.09 Project number 2733270 Page 2 of 2



The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Reports:

- No. NMi-2733270-01 dated 25 February 2022 that includes 51 pages;
- No. NMi-2733270-04 dated 25 February 2022 that includes 17 pages.

Characteristics of the non-automatic weighing instrument:

| Accuracy class | | | | III | | |
|-----------------------------------|----------------|--|--|---|--|--|
| Maximum capacity | | 30 kg ≤ Max ≤ 150 kg | | | | |
| Verification scale interval | | | | e ≥ 5 g | | |
| Weighing range | | Single interval | | | | |
| Maximum number of scale intervals | | | | n ≤ 6000 | | |
| Temperature range | | -10 °C / +40 °C | | | | |
| Power supply voltage | | 6,5 – 7,5 V DC supplied by any mains adapter or over the interface | | | | |
| Application | | Intended to be used for direct sales to the public | | | | |
| Software identification | Version number | (xx = | | u01.xx . 99 and represents the non-legally levant part of the software) | | |









