

**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-22.10 revision 0  
Project number 2495717  
Page 1 of 2

Issuing authority

NMi Certin B.V.  
Person responsible: M.Ph.D. Schmidt

**Applicant and  
Manufacturer**

Dräger Medical Systems, Inc.  
3135 Quarry Road  
Telford, PA 18969  
United States of America

Identification of the  
certified type

A **Non-automatic weighing instrument**  
Type : TN300

Characteristics

See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 2006 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority

**NMi Certin B.V., OIML Issuing Authority NL1**  
12 May 2022

Certification Board

NMi Certin B.V.  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 6362332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.



**OIML Member State**  
The Netherlands

Number R76/2006-A-NL1-22.10 revision 0  
Project number 2495717  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Type Evaluation Report:

- No. NMI-2495717-01 dated 26 April 2022 that includes 58 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	III
Maximum capacity	Max $\leq$ 15 kg
Verification scale interval	$e \geq$ 10 g
Weighing range(s)	Single interval
Maximum number of scale intervals	$n \leq$ 1500
Tare	$T \leq$ -Max
Temperature range	15 °C / 45 °C
Power supply voltage	100 – 240 V AC 50/60 Hz
Application	To be used only for weighing patients
Software identification	Version number
	5.01

## Revision History

This revision replaces the previous version.

Revision	Date	Changes
Initial	2022-05-12	-