



OIML Member State

Czech Republic

OIML Certificate No.

R85/2008-B-CZ1-2020.01

OIML CERTIFICATE ISSUED UNDER SCHEME B

OIML Issuing Authority

Name: **Czech Metrology Institute**

Address: Okružní 31

638 00 Brno

Czech Republic

Person responsible: Jan Kalandra

Applicant

Name: **Start Italiana SRL**

Address: Via Pola 6

Bovisio Masciago (MB)

Italy

Manufacturer

Name: **Start Italiana SRL**

Address: Via Pola 6

Bovisio Masciago (MB)

Italy

Identification of the certified type (*the detailed characteristics will be defined in the additional pages*)

Magnetostrictive level gauge

type: DMP-IS-485

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 85, Edition: 2008



This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report and test reports:

No. 015_19 that includes 1 page,

No. 6015-PT-P5002-20 that includes 29 pages, No. 6015-PT-P3016-14 that includes 30 pages and No. 6015-PT-P3024-12 that includes 34 pages.

The technical documentation relating to the identified type is contained in documentation file:

015_19 dated 11 October 2019.

OIML Certificate History

Revision No.	Date	Description of the modification

The OIML Issuing Authority

RNDr. Pavel Klenovský

Director General

Date: 29 June 2020



Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

Measuring system description

The level sensor DMP-IS-485 is done by a shaft in stainless steel, by a head in stainless steel and composite caps IP68, a circular connection 7/8" or cable gland and fuel resistant cable, floats in Stainless Steel, or NBR, or in special expanded cell close PVC.

The level sensor can be used with consoles MAGLINK-LX, SiteSentinel Nano and SiteSentinel Integra 100/500 as indicating unit.

Characteristics

Probes	Type	DMP-IS-485 (option: 1-wire)
	Measuring principle	Magnetostrictive
	Probe's body material	Stainless steel
	Floats	Stainless steel, NBR or expanded PVC
	Length	up to 7000 mm
	Accuracy	better than ± 1 mm
	Temperature	- 25 °C to + 55 °C