

OIML Member state
The Netherlands

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1,
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Endress+Hauser GmbH+Co., KG
Address: Hauptstrasse 1
Maulburg
Germany

Manufacturer of the certified type

Name: Endress+Hauser
Address: Hauptstrasse 1
Maulburg
Germany

Identification of the certified type

Automatic level gauge for measuring the level of liquid in storage tanks, model FMR 540,
for free space applications, with horn antenna DN100 or with parabolic antenna DN200.

NMI Certin BV
Hugo de Grootplein 1, 3314 EG Dordrecht
P.O. Box 394, 3300 AJ Dordrecht, NL
phone +31 78 6332332
fax +31 78 6332309
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

The notification of NMI Certin as Issuing Authority can be verified at www.oiml.org.



Specifications

Minimum product level	Maximum product level	Minimum and maximum for liquid pressure, temperature and properties	Minimum and maximum for vapour pressure, temperature and properties
23 m below antenna	0,1 m below antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.
26 m below antenna	0,1 m below antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R85
Edition 2008 (E)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

- R85/2008-NL1-09.01 that includes 66 pages
- R85/2008-NL1-09.02 that includes 19 pages

The Issuing Authority NL1
NMI Certin, 30 June 2009


C. Oosterman
Head Certification Board

This revision 1 replaces the previous revision, which was named "Certificate N° R85/2008-NL1-09.01".

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.