

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1,
3314 EG Dordrecht
The Netherlands

Applicant

Name: Enraf B.V.
Address: Delftechpark 39
2628 XJ Delft
The Netherlands

Manufacturer of the certified type

Name: Enraf B.V.
Address: Delftechpark 39
2628 XJ Delft
The Netherlands

Identification of certified type

Automatic level gauge for measuring the level of liquid in storage tanks, models Smartradar Flexline XP and Smartradar Flexline HP, with the antennas F08, W06, H04, S06, S08, S10 and S12.

Antenna type	Minimum product level	Maximum product level	Minimum and maximum values for liquid pressure, for liquid temperature and for liquid properties.	Minimum and maximum values for vapour pressure, for vapour temperature and for vapour properties.
F08, W06	20 m below the antenna	1 m below the antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.
H04, S06, S08, S10, S12	21 m below the antenna	1 m below the antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.



OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R85

Edition 2008 (E)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

R85/1998-NL1-07.02 that includes 100 pages;

CPC/9200376 that includes 20 pages;

NMi-12200691 that includes 13 pages.

NMi Certin B.V., OIML Issuing Authority NL1

7 December 2012

C. Oosterman

Head Certification Board

*

* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.