

OIML Member State
The Netherlands

Number R85/2008-NL1-17.04 revision 3
Project number 1901159
Page 1 of 2

Issuing authority Person responsible: NMI Certin B.V.
C. Oosterman

Applicant and Manufacturer Endress+Hauser Yamanashi Co., Ltd.
862-1 Mitsukunugi Sakaigawa-cho
Fuefuki-shi Yamanashi 406-0846
Japan

Identification of the certified type An **automatic level gauge**
Type: NMS8x

Characteristics See page 2 and further

This Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML Type Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 85-1 & 2 (2008) "Automatic level gauges for measuring the level of liquid in stationary storage tanks"

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**
28 August 2017



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at www.oiml.org



OIML Member State
The Netherlands

Number R85/2008-NL1-17.04 revision 3
Project number 1901159
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated report(s):

- No. NMI-16200591-03 dated 19 December 2016 that includes 47 pages.

Characteristics of the measuring instrument

In Table 1 the general characteristics of the measuring instrument are presented.
The construction of the measuring instrument is recorded in the Documentation folder no. TC8911-1.

Table 1 General characteristics

Measuring range	See table 2
Ambient temperature range	-40 – +70 °C; condensing humidity
Power supply voltage	100 - 240Vac (-15% / +10%) @ 50/60Hz
Software identification	Version number: 01.02.00 with checksum: 0x477A. Version number: 01.02.01 with checksum: 0xB028.

The complete family of measuring instruments (which are of similar construction) have the following characteristics indicated in table 2.

Table 2 General characteristics of the family of instruments

Type	Maximum measuring range
Type NMS8x	40 meter

Certificate history:

This revision replaces the previous revision.

Revision	Date	Description of the modification
Initial	24 April 2017	-
01	21 July 2017	New software version because of editorial changes.
02	26 July 2017	Editorial change.
03	28 August 2017	Typing error in date on certificate history