

# Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML  
Germany



OIML Certificate N°  
**R107/1997-DE1-05.02**

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Roman Schwartz

### Applicant

Name: BMT GmbH  
Address: Schäferstrasse 34a, 59174 Kamen  
GERMANY

Manufacturer of the certified type is the applicant.

### Identification of the certified type

Discontinuous Totalizing Automatic Weighing Instrument  
Type: Bulk Controller BMT

Accuracy class  $\geq 0.2$   
 $n \leq 6000$  ( $n \leq 3000$  per partial weighing range with multi-interval instruments)  
Max 3 kg to 300 t

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Report / Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R107-1**, edition 1997  
for accuracy class 0.2, 0.5; 1 and 2 (2 only for special purposes)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

# Physikalisch-Technische Bundesanstalt

OIML Certificate N°  
**R107/1997-DE1-05.02**

The conformity was established by the results of tests and examinations provided in the associated Report 1.12-4017747 (15 pages) and Test Report No. 1.12-4017747/1 (50 pages).

## The Issuing Authority

Dr. R. Schwartz  
Direktor und Professor

26.09.2005

## The OIML Member

Prof. Dr. M. Kochsiek  
Vizepräsident

26.09.2005

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.