

Nederlands Meetinstituut

**OIML Member state**  
The Netherlands

**OIML Certificate N° R107/1997-NL1-05.01**

Project number 502905

Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMi Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Penko Engineering B.V.  
Address: Wageningse laan 52-54  
3903 LA Veenendaal  
The Netherlands

### Manufacturer of the certified type

Name: Penko Engineering B.V.  
Address: Wageningse laan 52-54  
3903 LA Veenendaal  
The Netherlands

### Identification of certified type

Type : DisCon

Max  $\leq$  300 t

Min  $\geq$  20 g

$\Sigma$ min  $\geq$  1.5 kg

$d_t \geq$  1 g

$\Sigma$ min  $\geq$  400  $d_t$  and  $0.01\% \leq d_t \leq 0.1\%$  of Max

Temperature range -10 °C / +40 °C


**NMi Certin B.V.**  
Hugo de Grootplein 1, 3314 EG Dordrecht  
P.O. Box 394, 3300 AJ Dordrecht, NL  
phone +31 78 6332332  
fax +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi B.V. (see "Regulation objection and appeal against decisions of NMi B.V.")

NMi Certin B.V., chamber o.c. nr. 27.233.418

This document is issued under the provision that no responsibility is accepted and that the applicant gives warranty for each responsibility against third parties.

The notification of NMi Certin as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org).



Nederlands Meetinstituut

**OIML Member state**  
The Netherlands

**OIML Certificate N° R107/1997-NL1-05.01**

Project number 502905

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6752 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R107**

Edition 1997 (E)  
for accuracy class 0.5

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval. The conformity was established by the results of tests and examinations provided in the associated Test Reports:

DisCon:

- N° R107/1997-NL1-05.01 that includes 65 pages;

MassteQ:

- N° 201385 that includes 15 pages;

Minimate:

- N° 300513 that includes 23 pages;

QMA, Smart:

- N° 1017718 that includes 46 pages;
- N° 1017718 zener version that includes 26 pages;

QMA-ex, Smart-ex:

- N° 201356 that includes 18 pages;


AMI:

- N° 10041436 that includes 42 pages;
- N° 10041436B that includes 25 pages;
- N° 10049648 that includes 10 pages;
- N° 10058689 that includes 14 pages;

Load cell:

- N° 97008114/L that includes 36 pages;
- N° 3005349/L that includes 36 pages;
- N° 3016228/L that includes 36 pages.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification



23 September 2005

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.