



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Precia Molen
Address: Franse Akker 1
4824 AL Breda
The Netherlands

Manufacturer of the certified type

Name: Precia Molen
Address: Franse Akker 1
4824 AL Breda
The Netherlands

Identification of certified type

Discontinuous totalizing automatic weighing instrument
Type: ABS

Max, Min, Σ_{min} and d_t depend on the components used.

Climatic environment: -10 °C / +40 °C
85% humidity at 40 °C non-condensing

Electromagnetic environment class: E2



OIML Member state
The Netherlands

OIML Certificate N° R107/1997-NL1-09.01
Project number 806116
Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report, the type-approval certificate and the description with number T10162 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R107
Edition 1997 (E)
for accuracy class 0.2, 0.5, 1 and 2


This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R107/1997-NL1-09.01 that includes 43 pages;
N° 10107116 that includes 43 pages;
N° 1001943 that includes 79 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification


29 January 2009

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, though they may be reproduced in full.