

Member State
P.R.China

OIML Certificate No.
R117/1995-CN1-10.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

Applicant

Name: WenZhou Realtech Petroleum Equipment Limited Company.
Address: West 1/F, Building 1, No.108, Juguang Middle Road, Juguangyuan,
WenZhou City, Zhejiang, (325029) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Characteristics: Fuel dispenser

Model designation	RT-C 112	RT-C 122	RT-C 124	RT-C 222	RT-C 224
Accuracy class	0.5				
Maximum flowrate	50 L/min				
Minimum flowrate	5 L/min				
Minimum Measured Quantity	5 L				
Liquid	Gasoline/diesel oil/kerosene				
Maximum unit price (number of digits)	4				
Maximum price to pay (number of digits)	6				
Voltage	220V/380V				
Operating Temperature range	(-25~55)°C				
RT-C 112	1 product, 1 hose, 2 display with pump		RT-C 122	1 product, 2 hose, 2 display with pump	
RT-C 124	1 product, 2 hose, 4 display with pump		RT-C 222	2 product, 2 hose, 2 display with pump	
RT-C 224	2 product, 2 hose, 4 display with pump		/		

Comprising of: (type of RT-C 224)

- Two JBL-50 combined pumps and gas eliminator devices
- Two YBB3341EA motors for 220V/0.75kW
- Two BLS-22KQ-3NM electronic controller devices
- Two DSF-20 solenoid valves
- Two J100 flow meters
- Two B-TY sensors

The numbers of the associated test reports and the included pages: 2010 OM 001 52 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: R117 & R118

Edition: 1995

For accuracy class (if applicable): 0.5

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority



The CIML member

A handwritten signature in black ink, consisting of three stylized Chinese characters: '蒲长城' (Pu Changcheng).

June 10, 2010

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.